


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2022, Rack A, Plate 1, Well D12, 482099 FA
 [C93295C2DF0A8C4A] Dec 19, 2025 11:42:56 Pressure = 75 bar (74 to 77)



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.412	F 0.48	1.1%	19769	
2	0.500	F 0.58	0.6%	11867	
3	0.702	F 0.82	18.8%	350330	Acetylated F peak
4	0.842	F 0.98	67.8%	1261295	Consistent with F
5	1.228	A 1.01	11.6%	216284	A peak
Total Area:				1859545	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range