


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2073, Rack A, Plate 1, Well D07, 492239
 [D877610E612E2642] Feb 13, 2026 13:58:42 Pressure = 52 bar (52 to 53)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.410	F 0.46	1.3%	19151	
2	0.518	F 0.58	0.4%	6289	
3	0.753	F 0.85	17.0%	245837	Acetylated F peak
4	0.867	F 0.98	64.6%	937055	Consistent with F
5	1.022	F 1.15	0.5%	7424	
6	1.262	A 1.01	16.1%	234050	A peak
Total Area:				1449806	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range