


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2068, Rack A, Plate 1, Well C11, 498922
 [D973206E6CF35BB5] Feb 10, 2026 11:34:23 Pressure = 94 bar (94 to 97)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.425	F 0.50	1.1%	13335	
2	0.497	F 0.59	0.4%	5221	
3	0.833	F 0.99	72.8%	874568	Consistent with F
4	1.008	F 1.20	2.6%	31403	
5	1.205	A 1.02	22.8%	273731	A peak - REVIEW
6	1.390	S 0.89	0.2%	2886	
Total Area:				1201144	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range