


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1993, Rack A, Plate 1, Well B11, 171165
[C922BD92FD328640] Nov 25, 2025 11:17:39

Pressure = 70 bar (70 to 72)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.395	F	0.46	1.5%	11797	
2	0.497	F	0.58	1.2%	9600	
3	0.718	F	0.84	16.0%	130219	Acetylated F peak
4	0.870	F	1.02	54.5%	442724	Consistent with F
5	1.000	F	1.17	2.0%	16528	
6	1.233	A	1.01	24.8%	201635	A peak - REVIEW
Total Area:					812503	

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range