


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2110, Rack A, Plate 1, Well G05, 172882  
[8D774414092BF9BF] Mar 31, 2026 13:39:24 Pressure = 58 bar (58 to 60)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.515	F 0.58	2.3%	43746	
2	0.723	F 0.81	16.6%	315827	Acetylated F peak
3	0.858	F 0.96	67.2%	1278913	Consistent with F
4	1.248	A 1.01	14.0%	265715	A peak
Total Area: 1904201					

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range