


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2049, Rack A, Plate 1, Well B04, 488989
 [CD678D5F96E8266A] Jan 20, 2026 11:29:24 Pressure = 80 bar (80 to 82)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.507	F 0.60	1.2%	12965	
2	0.700	F 0.83	16.4%	180956	Acetylated F peak
3	0.830	F 0.99	63.0%	695187	Consistent with F
4	0.987	F 1.17	1.2%	13029	
5	1.198	A 1.01	18.2%	200736	A peak
Total Area:				1102873	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range