


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1860, Rack A, Plate 1, Well B10, 458027  
 [8C30885DCBBD06C0] Jul 09, 2025 10:39:23 Pressure = 61 bar (61 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.400	F 0.47	1.7%	16371	
2	0.487	F 0.57	3.4%	33764	
3	0.703	F 0.83	9.4%	91955	Acetylated F peak
4	0.845	F 1.00	64.6%	633193	Consistent with F
5	1.232	A 1.02	20.9%	204363	A peak - REVIEW
Total Area:				979646	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range