


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1833, Rack A, Plate 1, Well E01, 785487
[C93187B8B280D91D] Jun 18, 2025 11:38:41 Pressure = 60 bar (59 to 61)

FA



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|-------------|-------|--------|--------|---------|-------------------|
| 1 | 0.493 | F 0.58 | 7.5% | 160791 | |
| 2 | 0.835 | F 0.98 | 84.1% | 1803733 | Consistent with F |
| 3 | 1.228 | A 1.02 | 8.4% | 180226 | A peak |
| Total Area: | | | | 2144750 | |

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range