


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1930, Rack A, Plate 1, Well C04, 170435
[D960524FB42B84C8] Oct 03, 2025 11:07:21 Pressure = 62 bar (61 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.405	F 0.48	1.4%	11415	
2	0.488	F 0.58	0.9%	7389	
3	0.697	F 0.83	11.9%	98803	Acetylated F peak
4	0.842	F 1.01	58.2%	481477	Consistent with F
5	0.973	F 1.16	1.9%	16034	
6	1.195	A 1.00	25.7%	212414	A peak - REVIEW
Total Area:				827532	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range