


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2176, Rack A, Plate 1, Well D06, 493520
[D87CA9142BBA6CC2] Jun 12, 2026 13:13:46 Pressure = 68 bar (68 to 70)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.515	F 0.61	1.7%	18617	
2	0.703	F 0.83	15.3%	163505	Acetylated F peak
3	0.848	F 1.01	74.2%	791742	Consistent with F
4	1.187	A 1.00	8.7%	93066	A peak
Total Area:				1066930	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range