


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2197, Rack A, Plate 1, Well B09, 503347
[CC7CD6CAEA88E4E2] Jul 15, 2026 12:57:18 Pressure = 83 bar (83 to 85)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.432	F	0.52	0.9%	3939	
2	0.518	F	0.63	0.5%	2056	
3	0.680	F	0.82	16.2%	71443	Acetylated F peak
4	0.837	F	1.01	70.3%	311032	Consistent with F
5	1.170	A	1.01	12.1%	53663	A peak
Total Area:					442133	

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range