


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2077, Rack A, Plate 1, Well D06, 489343  
[986223D2F77BD39F] Feb 18, 2026 11:11:20 Pressure = 56 bar (55 to 57)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.422	F 0.48	1.0%	13827	
2	0.515	F 0.59	1.1%	15051	
3	0.730	F 0.83	20.0%	278312	Acetylated F peak
4	0.878	F 1.00	70.3%	979718	Consistent with F
5	1.257	A 1.00	7.7%	106642	A peak
Total Area:				1393550	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range