


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2061, Rack A, Plate 1, Well F03, 490389
 [D9778E05F63759B5] Jan 30, 2026 12:10:55 Pressure = 89 bar (88 to 92)

AF



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.403	F 0.48	1.2%	9745	
2	0.492	F 0.58	0.5%	4099	
3	0.693	F 0.82	4.4%	35566	Acetylated F peak
4	0.837	F 0.99	33.1%	264955	Consistent with F
5	0.995	F 1.18	5.7%	45401	3
6	1.198	A 1.00	54.4%	435429	A peak - REVIEW
7	1.382	S 0.89	0.7%	5748	
Total Area:				800943	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range