


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1364, Rack A, Plate 1, Well B04, 65047
[D8DDF6800A25373D] Apr 08, 2024 10:04:05 Pressure = 61 bar (61 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.485	F 0.55	4.1%	59626	
2	0.737	F 0.83	12.7%	182256	Acetylated F peak
3	0.868	F 0.98	68.4%	982837	Consistent with F
4	1.262	A 1.02	14.8%	212530	A peak
Total Area:				1437249	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range