


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2194, Rack A, Plate 1, Well A06, 487359
[8C69D15AC1A26EE0] Jul 09, 2026 10:29:19 Pressure = 81 bar (81 to 83)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.518	F	0.63	1.1%	10010	
2	0.688	F	0.83	15.1%	134394	Acetylated F peak
3	0.818	F	0.99	75.4%	671374	Consistent with F
4	1.170	A	1.02	8.3%	74323	A peak
Total Area: 890101						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range