


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2197, Rack A, Plate 1, Well H05, 486050
[9C7D96DFB7D5931F] Jul 15, 2026 15:20:48

Pressure = 83 bar (83 to 85)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.687	F	0.82	18.5%	271147	Acetylated F peak
2	0.820	F	0.98	72.2%	1056733	Consistent with F
3	1.155	A	1.00	9.3%	135535	A peak
Total Area: 1463415						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range