


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2143, Rack A, Plate 1, Well E06, 496533
 [99360F094361D3BF] May 08, 2026 13:13:00 Pressure = 63 bar (63 to 65)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.428	F 0.50	0.9%	9133	
2	0.512	F 0.59	0.5%	5236	
3	0.702	F 0.82	13.6%	137570	Acetylated F peak
4	0.848	F 0.99	67.8%	687521	Consistent with F
5	1.222	A 1.02	17.2%	174118	A peak
Total Area:				1013578	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range