


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2204, Rack A, Plate 1, Well G06, 500830
[8868D9E96D3F3137] Jul 23, 2026 12:50:46 Pressure = 94 bar (94 to 96)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.513	F	0.61	3.8%	42202	
2	0.822	F	0.97	76.2%	856840	Consistent with F
3	1.158	A	1.00	20.1%	225792	A peak - REVIEW
Total Area: 1124834						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range