


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2149, Rack A, Plate 1, Well F10, 501555
[9828A4494B85BB1D] May 20, 2026 14:30:09 Pressure = 64 bar (64 to 66)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.433	F	0.50	0.9%	9394	
2	0.518	F	0.60	0.5%	5764	
3	0.712	F	0.82	18.8%	204831	Acetylated F peak
4	0.847	F	0.98	75.2%	817204	Consistent with F 6
5	1.218	A	1.01	4.6%	49888	A peak
Total Area: 1087081						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range