


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2029, Rack A, Plate 1, Well A06, 171659
 [CD77C7FDC7775B97] Dec 29, 2025 11:02:18 Pressure = 76 bar (76 to 78)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.410	F 0.48	1.0%	10188	
2	0.497	F 0.59	0.7%	6828	
3	0.753	F 0.89	16.4%	169837	Acetylated F peak
4	0.838	F 0.99	70.0%	722764	Consistent with F
5	1.005	F 1.19	0.7%	7088	
6	1.223	A 1.02	11.3%	116218	A peak
Total Area:				1032923	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range