


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2185, Rack A, Plate 1, Well E12, 486028
[8D39BB56749A6C48] Jun 25, 2026 13:49:53

FA
Pressure = 72 bar (72 to 74)



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.427	F	0.50	0.3%	3209	
2	0.522	F	0.61	0.3%	2519	
3	0.678	F	0.80	13.8%	134467	Acetylated F peak
4	0.827	F	0.97	74.9%	728685	Consistent with F
5	1.195	A	1.01	10.7%	103703	A peak
Total Area:					972583	

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range