


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2179, Rack A, Plate 1, Well A01, 486155
[893CAB8402BA4E62] Jun 18, 2026 12:33:58 Pressure = 70 bar (70 to 72)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.515	F 0.62	2.2%	13577	
2	0.695	F 0.83	17.2%	106316	Acetylated F peak
3	0.838	F 1.00	70.1%	434196	Consistent with F
4	1.192	A 1.00	10.5%	65267	A peak

Total Area: 619356

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range