


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2194, Rack A, Plate 1, Well D01, 493070
[8D6CC44AEBA86460] Jul 09, 2026 11:33:59 Pressure = 81 bar (81 to 83)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.522	F	0.63	1.1%	10939	
2	0.687	F	0.83	10.1%	100995	Acetylated F peak
3	0.823	F	0.99	63.6%	635980	Consistent with F
4	1.150	A	1.00	25.3%	252798	A peak - REVIEW
Total Area: 1000712						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range