


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2061, Rack A, Plate 1, Well C12, 489186
 [D832CA45FC9FF995] Jan 30, 2026 11:14:19 Pressure = 89 bar (89 to 92)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.415	F 0.49	0.7%	10852	
2	0.502	F 0.60	0.6%	9188	
3	0.690	F 0.82	15.2%	240895	Acetylated F peak
4	0.827	F 0.98	59.3%	940172	Consistent with F
5	0.983	F 1.17	1.4%	22352	
6	1.195	A 1.01	22.8%	362022	A peak - REVIEW
Total Area:				1585481	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range