


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2054, Rack A, Plate 1, Well B02, 483339
 [D8769ECAC81FF117] Jan 26, 2026 11:06:10 Pressure = 88 bar (87 to 90)

FA



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|-------------|-------|--------|--------|--------|-------------------|
| 1 | 0.408 | F 0.49 | 1.4% | 9880 | |
| 2 | 0.497 | F 0.59 | 0.5% | 3355 | |
| 3 | 0.688 | F 0.82 | 13.5% | 95310 | Acetylated F peak |
| 4 | 0.833 | F 1.00 | 63.7% | 450232 | Consistent with F |
| 5 | 0.993 | F 1.19 | 2.5% | 17506 | |
| 6 | 1.203 | A 1.02 | 18.5% | 130591 | A peak |
| Total Area: | | | | 706874 | |

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range