


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2124, Rack A, Plate 1, Well D01, 502205
[DC2358E30FBC26E8] Apr 16, 2026 11:25:17 Pressure = 62 bar (62 to 64)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.515	F 0.59	3.3%	33856	
2	0.725	F 0.82	17.6%	181671	Acetylated F peak
3	0.882	F 1.00	67.4%	696324	Consistent with F
4	1.237	A 1.00	11.7%	121134	A peak
Total Area:				1032985	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range