


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1944, Rack A, Plate 1, Well C09, 479963
 [CC614D36513E51BF] Oct 10, 2025 11:28:44 Pressure = 67 bar (67 to 68)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.412	F 0.49	0.9%	9669	
2	0.490	F 0.58	0.7%	7100	
3	0.678	F 0.81	10.9%	111752	Acetylated F peak
4	0.830	F 0.99	57.7%	593043	Consistent with F
5	0.980	F 1.17	2.3%	23366	
6	1.198	A 1.00	27.0%	277689	A peak - REVIEW
7	1.362	A 1.14	0.6%	5898	
Total Area:				1028517	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range