


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2179, Rack A, Plate 1, Well G09, 485915  
[9D7CBA942812ECE2] Jun 18, 2026 15:19:42 Pressure = 70 bar (70 to 72)

FA



| PEAK                | RT    | REL | RT   | % CONC | AREA   | COMMENT           |
|---------------------|-------|-----|------|--------|--------|-------------------|
| 1                   | 0.428 | F   | 0.51 | 0.9%   | 11506  |                   |
| 2                   | 0.507 | F   | 0.60 | 0.3%   | 4638   |                   |
| 3                   | 0.712 | F   | 0.85 | 13.0%  | 174608 | Acetylated F peak |
| 4                   | 0.825 | F   | 0.98 | 68.0%  | 916075 | Consistent with F |
| 5                   | 1.198 | A   | 1.01 | 17.8%  | 240148 | A peak            |
| Total Area: 1346975 |       |     |      |        |        |                   |

Codes: 1) Wide A peak  
2) Area of A peak < 80%  
3) Peak area greater than expected  
4) Peak after A2  
5) A1c > 10%  
6) HbF or variant present  
7) Total sample area too small/big  
8) A2 is not within normal range