


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2149, Rack A, Plate 1, Well C09, 501892
[8978B152F4D8EE48] May 20, 2026 13:13:21 Pressure = 65 bar (65 to 67)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.507	F 0.59	3.0%	34600	
2	0.710	F 0.83	11.5%	131145	Acetylated F peak
3	0.848	F 0.99	67.6%	767717	Consistent with F
4	1.220	A 1.01	17.8%	202791	A peak
Total Area:				1136253	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range