


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2200, Rack A, Plate 1, Well D06, 493014
[CD3C976DCC006C40] Jul 17, 2026 15:40:51 Pressure = 82 bar (82 to 85)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.525	F	0.63	0.8%	13581	
2	0.692	F	0.83	11.0%	183643	Acetylated F peak
3	0.820	F	0.98	61.6%	1030406	Consistent with F
4	1.145	A	1.00	26.6%	445078	A peak - REVIEW
Total Area: 1672708						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range