


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2161, Rack A, Plate 1, Well G02, 501523
[DD29A367ACAD33B5] May 29, 2026 13:39:43

Pressure = 61 bar (61 to 63)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.427	F	0.50	0.6%	6524	
2	0.508	F	0.60	0.6%	6971	
3	0.705	F	0.83	13.3%	150989	Acetylated F peak
4	0.845	F	1.00	66.0%	751426	Consistent with F
5	1.197	A	0.99	19.6%	222658	A peak
Total Area: 1138568						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range