


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1993, Rack A, Plate 1, Well A02, 482223  
 [8D32A9C7AAEFD11D] Nov 25, 2025 10:35:13 Pressure = 70 bar (70 to 72)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.400	F 0.47	1.2%	11052	
2	0.497	F 0.58	0.8%	6944	
3	0.717	F 0.84	17.4%	156202	Acetylated F peak
4	0.870	F 1.02	66.0%	590754	Consistent with F
5	1.235	A 1.01	14.6%	130509	A peak
Total Area:				895461	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range