


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1833, Rack A, Plate 1, Well A03, 456872
[8D24C3F8B8287BBD] Jun 18, 2025 10:03:41 Pressure = 61 bar (60 to 62)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.508	F 0.60	2.9%	18088	
2	0.700	F 0.83	16.3%	103185	Acetylated F peak
3	0.837	F 0.99	70.1%	444784	Consistent with F
4	0.978	F 1.16	2.0%	12858	
5	1.225	A 1.01	8.7%	55380	A peak
Total Area:				634295	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range