


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2194, Rack A, Plate 1, Well G11, 493094
[C92CD50AE300C4C0] Jul 09, 2026 13:08:59

Pressure = 79 bar (79 to 82)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.440	F	0.53	1.4%	18174	
2	0.505	F	0.61	1.1%	13285	
3	0.690	F	0.83	13.9%	175012	Acetylated F peak
4	0.817	F	0.98	72.0%	905658	Consistent with F
5	1.155	A	1.00	11.6%	145350	A peak
Total Area: 1257479						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range