


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2147, Rack A, Plate 1, Well D01, 173354
 [C978F4D4A8FAE4EA] May 13, 2026 15:20:12 Pressure = 67 bar (64 to 68)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.428	F 0.23	1.7%	25604	
2	0.560	F 0.30	4.4%	64840	
3	0.867	F 0.46	23.3%	344894	
4	1.300	F 0.70	0.4%	6035	
5	1.888	F 1.01	43.6%	644570	Consistent with F
6	2.620	F 1.40	4.7%	69910	
7	2.897	A 0.79	0.5%	7458	
8	3.208	A 0.87	0.3%	4022	
9	3.665	A 1.00	20.6%	304616	A peak - REVIEW
10	4.207	A 1.14	0.2%	2277	
11	4.407	A 1.20	0.2%	3196	
				Total Area: 1477422	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range