


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1359, Rack A, Plate 1, Well F05, 27442
 [8C8CA4038B25B71D] Apr 03, 2024 11:46:25 Pressure = 59 bar (59 to 60)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.418	F 0.47	1.5%	27817	
2	0.487	F 0.55	2.9%	53410	
3	0.762	F 0.86	9.8%	180285	Acetylated F peak
4	0.873	F 0.98	76.6%	1401765	Consistent with F
5	1.247	A 1.00	9.1%	167270	A peak
Total Area:				1830547	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range