


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2068, Rack A, Plate 1, Well C09, 498920
[8926353B39AC0CEA] Feb 10, 2026 11:30:20 Pressure = 95 bar (94 to 97)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.408	F 0.48	1.1%	7471	
2	0.500	F 0.59	0.5%	3433	
3	0.688	F 0.82	15.1%	103142	Acetylated F peak
4	0.832	F 0.99	59.7%	408724	Consistent with F
5	0.993	F 1.18	2.2%	15062	
6	1.197	A 1.01	21.4%	146327	A peak - REVIEW
Total Area:				684159	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range