


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2080, Rack A, Plate 1, Well E12, 489572
 [DC336247A9F17395] Feb 24, 2026 12:27:07 Pressure = 56 bar (56 to 58)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.417	F 0.47	1.2%	15835	
2	0.512	F 0.58	0.8%	11467	
3	0.732	F 0.83	14.5%	198634	Acetylated F peak
4	0.873	F 0.99	71.5%	980958	Consistent with F
5	1.267	A 1.01	12.0%	164808	A peak
Total Area:				1371702	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range