


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1977, Rack A, Plate 1, Well D07, 478845
 [9C36E6DB3DD28C68] Nov 14, 2025 11:36:37 Pressure = 68 bar (68 to 70)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.405	F 0.47	0.8%	10463	
2	0.498	F 0.58	0.7%	9169	
3	0.720	F 0.84	15.1%	188229	Acetylated F peak
4	0.852	F 0.99	69.0%	859692	Consistent with F
5	1.237	A 1.01	14.3%	177881	A peak
Total Area:				1245434	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range