


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1988, Rack A, Plate 1, Well C08, 171127
[D972E8E08D328E62] Nov 21, 2025 11:42:46 Pressure = 69 bar (68 to 71)

FA



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|-------------|-------|--------|--------|--------|-------------------|
| 1 | 0.400 | F 0.47 | 1.7% | 16105 | |
| 2 | 0.488 | F 0.57 | 1.6% | 15339 | |
| 3 | 0.708 | F 0.83 | 10.6% | 100147 | Acetylated F peak |
| 4 | 0.852 | F 0.99 | 56.4% | 532763 | Consistent with F |
| 5 | 1.003 | F 1.17 | 1.2% | 11400 | |
| 6 | 1.235 | A 1.01 | 28.4% | 268408 | A peak - REVIEW |
| Total Area: | | | | 944162 | |

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range