


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2140, Rack A, Plate 1, Well A03, 503265
[DC734E7B33EBF115] May 06, 2026 11:31:47 Pressure = 65 bar (64 to 67)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.517	F 0.61	2.4%	16941	
2	0.703	F 0.82	16.8%	120701	Acetylated F peak
3	0.848	F 0.99	65.3%	468937	Consistent with F
4	0.985	F 1.15	3.3%	23640	
5	1.232	A 1.01	12.3%	88209	A peak
Total Area:				718428	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range