


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1998, Rack A, Plate 1, Well E05, 484312
[8822E932F9B82660] Nov 28, 2025 12:35:35 Pressure = 71 bar (70 to 72)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.392	F 0.45	2.9%	29450	
2	0.492	F 0.57	0.9%	9503	
3	0.715	F 0.83	12.0%	122169	Acetylated F peak
4	0.867	F 1.00	76.4%	775225	Consistent with F
5	1.253	A 1.02	7.8%	78827	A peak
Total Area:				1015174	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range