


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2003, Rack A, Plate 1, Well C09, 483940
[8932EFF17218ACEA] Dec 03, 2025 11:14:20 Pressure = 74 bar (73 to 75)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.47	1.4%	16586	
2	0.498	F 0.58	0.5%	5963	
3	0.707	F 0.82	11.8%	139666	Acetylated F peak
4	0.848	F 0.98	79.6%	939137	Consistent with F
5	1.242	A 1.02	6.7%	79092	A peak
Total Area:				1180444	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range