


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2092, Rack A, Plate 1, Well A04, 498463
[DC237AB78DCE8C60] Mar 11, 2026 11:30:15 Pressure = 57 bar (57 to 59)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.417	F	0.48	0.8%	9443	
2	0.517	F	0.59	0.6%	6967	
3	0.722	F	0.83	12.4%	143528	Acetylated F peak
4	0.862	F	0.99	67.4%	778478	Consistent with F
5	1.258	A	1.01	18.8%	217389	A peak
Total Area: 1155805						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range