


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2127, Rack A, Plate 1, Well C10, 498272
[D937589F9DBC2468] Apr 21, 2026 10:53:25 Pressure = 63 bar (63 to 65)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.415	F 0.48	1.7%	11852	
2	0.503	F 0.58	2.9%	20090	
3	0.712	F 0.82	11.5%	80837	Acetylated F peak
4	0.877	F 1.01	49.6%	347822	Consistent with F
5	1.237	A 1.00	34.3%	240522	A peak - REVIEW
Total Area:				701123	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range