


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2065, Rack A, Plate 1, Well F11, 490395
[987730EACD8E0C48] Feb 04, 2026 14:03:45 Pressure = 85 bar (84 to 87)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.405	F 0.48	1.0%	9635	
2	0.507	F 0.60	0.2%	1663	
3	0.682	F 0.81	13.0%	131723	Acetylated F peak
4	0.825	F 0.98	74.0%	749138	Consistent with F
5	1.193	A 1.00	11.9%	120045	A peak

Total Area: 1012204

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range