


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2204, Rack A, Plate 1, Well A09, 500813
[89799DF94517B93D] Jul 23, 2026 10:27:17

Pressure = 92 bar (92 to 94)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.518	F	0.63	4.9%	44773	
2	0.827	F	1.00	85.9%	782820	Consistent with F
3	1.163	A	1.01	9.1%	83327	A peak
Total Area: 910920						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range