


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2210, Rack A, Plate 1, Well A12, 487080
[CC68D87E66EA44CA] Jul 30, 2026 10:22:32 Pressure = 79 bar (79 to 81)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.418	F	0.51	1.2%	17176	
2	0.502	F	0.61	0.5%	6971	
3	0.810	F	0.98	80.6%	1165111	Consistent with F
4	1.190	A	1.00	17.7%	256142	A peak
Total Area: 1445400						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range