


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1924, Rack A, Plate 1, Well C06, 478894 FA
 [DC3157067C89A6E0] Oct 01, 2025 11:42:23 Pressure = 62 bar (62 to 63)



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.403	F 0.48	1.2%	9457	
2	0.495	F 0.59	0.9%	7237	
3	0.677	F 0.80	13.0%	104668	
4	0.825	F 0.98	61.6%	494020	Acetylated F peak Consistent with F
5	0.973	F 1.16	2.2%	17994	
6	1.198	A 1.01	20.7%	166054	A peak - REVIEW
7	1.343	A 1.13	0.4%	3132	
Total Area:				802562	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range