


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1988, Rack A, Plate 1, Well D10, 482079
 [9866FCE08DB80660] Nov 21, 2025 12:11:03 Pressure = 68 bar (68 to 70)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.47	1.2%	14122	
2	0.497	F 0.58	1.0%	10977	
3	0.720	F 0.84	14.9%	168011	Acetylated F peak
4	0.868	F 1.01	58.6%	662600	Consistent with F
5	1.003	F 1.17	1.6%	17672	
6	1.230	A 1.00	22.8%	257674	A peak - REVIEW
Total Area:				1131056	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range